



Certificate of Traceability

Manufacturer Certification

This certifies that Abbott Diagnostics, using HbA1c (Whole Blood) on Architect c8000 has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: Architect c8000	Calibrator Lot: 60939UQ11, 60939UQ11	Application: %NGSP=HbA1c/Hb*91.48
Reagent Lot: 61011UQ10, 61011UQ10, 61011UQ10	Calibrator Assigned Values: 5.1%, 11.4%	+2.152

Date of Certification: May 1, 2022

Certification Expires: May 1, 2023

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD

NGSP Network Coordinator

Sharon Amalf

SRL director/ supervisor